



**The US EPA recognizes AMCO Clear® as a Primary Standard for the calibration of turbidity meters used both online and in the laboratory for the analysis of drinking water.**

## FEATURES

- Safe, non-toxic and disposable
- Requires no mixing or inverting
- Easy-to-use / No dilutions or preparations
- Stable - does not settle out of suspension
- Accurate to 1% lot-to-lot
- Guaranteed one-year shelf life from date of shipment
- Available in a wide range of values
- N.I.S.T Traceable
- Custom standards available

Our Amco Clear® turbidity standards are traceable to NIST Primary Standards on account of their consistent particle size and photometric accuracy via UV-VIS spectroscopy. Each lot is normalized to historical lots which were formulated to dilutions of ISO17034 4000 NTU Formazin.



- EPA Method 180.1 Approved
- ISO 17025 Accredited



## TURBIDITY STANDARD KITS

**Primary Standard Turbidity Meter Calibration Kits** include 250 mL bottles of each NTU value at the manufacturer's suggested calibration points.

**Secondary Standard Turbidity Meter Kits** include sealed vials of the NTU values at the manufacturer's suggested calibration points.

**Deluxe Turbidity Standard Kits** will contain the 250 mL bottles of primary standards and the corresponding sealed secondary standards at the manufacturer's suggested calibration points. This will also include a new sample vial.

**Linear Calibration Standard Kits** contain 125 mL bottles of primary standard in the range of your application. 0-1, 0-10, and 0-100 NTU ranges include 6 bottles at 0%, 20%, 40%, 60%, 80% and 100% of the range of interest. Directions and graph paper are included upon request.

## INDIVIDUAL TURBIDITY STANDARDS

**Primary Standards** are available in 125, 250, and 500 mL, 1 L and 1 Gallon bottles and are instrument specific, formulated to match Formazin dilutions.

**Secondary Standards** are primary standard suspensions contained in a sealed vial/bottle and used for daily instrument calibration verification.

**Best Seller** is a low level check standard to verify instrument calibration in the drinking water range of interest.

